

BRIHANMUMBAI MUNICIPAL CORPORATION
CENTRAL PURCHASE DEPARTMENT

- Sub :** Minutes of Pre-bid meeting held on 11.07.2025 at 3.00 p.m. for
- 1) **SITC of Fully automated Immunochemistry Analyser (2 nos)** along with Standard Accessories with 3 years Warranty and 7 years CMC for KEM and Nair Hospitals.
 - 2) **SITC of Table Top Steam Autoclave with Distiller (13 nos.)** along with Standard Accessories with 3 years Warranty and 7 years CMC for Nair Hospital Dental College.
- Ref :**
1. Tender Id- 2025_MCGM_1196360_1.
 2. Tender Id- 2025_MCGM_1196365_1

Pre-bid meeting was arranged on **11.07.2025 at 3.00 p.m.** under the Chairmanship of A.M.C. (W.S.) to discuss & finalize the specifications of subject medical equipments to be procured for GTB Hospital, through e-tender. However, A.M.C. (W.S.) was pre-occupied in other meeting, the same was conveyed by D.M.C(CPD) when following officers of BMC were present.

Sr.No.	Name	Designation	Hospital/Department
1.	Smt. Sandhya Nandedkar	D.M.C(CPD)	CPD
2.	Dr. Neelam Andrade	Director (ME & MH)	NHDC
3.	Dr. Mohan Joshi	Dean (LTMG)	LTMG Hospital
4.	Dr. S.C. Mohite	Dean (Nair)	BYL Nair Hospital
5.	Dr. S.H. Ravat	Dean (KEM)	KEM Hospital
6.	Shri. Ajay Vasant Patil	DY.CH.E (CPD)	CPD
7.	Dr. Kusum Jashnani	HOD(Pathology)	BYL Nair Hospital
8.	Dr. Nitin Valvi	B.T.O. (Blood Centre)	BYL Nair Hospital
9.	Dr. Kavita Sawant	Addl Prof (Blood Centre)	LTMG Hospital
10.	Dr. Charu Agrawal	Lecturer (Pathology)	LTMG Hospital
11.	Dr. Jayashree Sharma	Prof & HOD (Transfusion medicin)	KEM Hospital
12.	Dr. K.S. Banga	HOD (Conservative Orthodontics)	NHDC Hospital
13.	Dr. Shweta Bhat	HOD (Orthodontics)	NHDC Hospital
14.	Shri. Vivian Kulkarni	E.E. (M&E) CPD i/c	CPD
15.	Smt. Anuja Lalpota	A.E (M &E) CPD i/c	CPD

1. **SITC of Fully automated Immunochemistry Analyser (2 nos)** along with Standard Accessories with 3 years Warranty and 7 years CMC for KEM and Nair Hospitals. Tender Id- 2025_MCGM_1196360_1.

Company representatives				
Sr. No.	Name	Name of Company	E-mail ID	Mobile no.

1.	Mr. Prathmesh Karkare	M/s. Bio-rad	Prathmesh_karkare@bio-rad.com	9757274229
2.	Mr.Sumit Kumar	M/s. Tulip Diagnostics	Sumit.kumar@tulipgroup.com	9967819131
3.	Mr.Urvish Shah	M/s. Crown Surgical	Shah.urvish@rediffmail.com	9820280930

The suggestions from bidder on point to point technical specifications were discussed. After pre-bid meeting, concerned HOD has made following changes.

Sr. No.	Specification of BMC	Suggestion of Bidder / Opinion of HOD	Decision taken
1	Latest model/version of fully automated equipment for ABO and Rh Typing, Rh phenotyping, Cross-match, antibody screening on gel cards based on Gel Column Agglutination Technology.	Latest model/version of fully automated equipment for ABO and Rh Typing, Rh phenotyping, extended phenotyping , Cross-match, antibody screening on gel cards based on Gel Column Agglutination Technology.	Latest model/version of fully automated equipment for ABO and Rh Typing, Rh phenotyping, extended phenotyping , Cross-match, antibody screening on gel cards based on Gel Column Agglutination Technology.
8	Should have a separate Transfer arm for transportation of gel cards between different modules. Separate Incubators for room temperature incubation and 37°C incubation. Each incubator should have minimum 24 gel cards capacity.	Should have a separate Transfer arm for transportation of gel cards between different modules. Separate Incubators for room temperature incubation and 37°C incubation. Each incubator should have minimum 24 gel cards capacity . 12*2 gel cards or more capacity	Should have a separate Transfer arm for transportation of gel cards between different modules. Separate Incubators for room temperature incubation and 37°C incubation. Each incubator should have minimum 24 gel cards capacity . 12*2 gel cards or more capacity
11	1 Centrifuge of 85g with 24 cards capacity.	1 Centrifuge of 85g with 24 cards capacity 12*2 gel cards or more capacity .	1 Centrifuge of 85g with 24 cards capacity 12*2 gel cards or more capacity .
12	Sample loading capacity should be more than 90 primary tubes.	Sample loading capacity should be more than 90 50 primary tubes.	Sample loading capacity should be more than 90 50 primary tubes.
13	Should have 1 dilution plate of 96 wells for preparation of red cell suspension.	Should have 1 dilution plate of 96 16 or more wells for preparation of red cell suspension.	Should have 1 dilution plate of 96 16 or more wells for preparation of red cell suspension.

2. **SITC of Table Top Steam Autoclave with Distiller (13 nos.)** along with Standard Accessories with 3 years Warranty and 7 years CMC for Nair Hospital Dental College. Tender Id- 2025_MCGM_1196360_1.

Company representatives				
Sr. No.	Name	Name of Company	E-mail ID	Mobile no.
1.	Mr. Jawed Alam	M/s. CL Micromed Pvt. Ltd.	jawed@clmicromed.com	9321213418
2.	Dinesh Kamde	M/s. My Healthskape	kamdedinesh@myhealthskape.com	9987096976
3.	Sital Shah	M/s. PEI&C	Sital_shah@yahoo.com	9820033597
4	Chinmay Karnik	M/s. Equitron	Chinmayk@mediainstrument.com	8291851549
5	Sanjay Kothari	M/s. Kalindi	Kalindi.kothari@gmail.com	9322244604

The suggestions from bidder on point to point technical specifications were discussed. After pre-bid meeting, concerned HOD has made following changes

Sr. No.	Specification of BMC	Suggestion of Bidder / Opinion of HOD	Decision taken
2	Jacketed inner chamber made of medical grade stainless steel (316) or equivalent grade capable of withstanding high temperature under pressure.	Jacketed inner chamber made of medical grade stainless steel (316) or equivalent higher grade capable of withstanding high temperature under pressure.	Jacketed inner chamber made of medical grade stainless steel (316) or equivalent higher grade capable of withstanding high temperature under pressure.
7	Should have minimum five precent sterilization programmes open & closed for metal instruments, liquids & delicate instruments and two test cycles one nos. each for sterilization and vaccum	Should have minimum five precent sterilization programmes open & closed for metal instruments, liquids & delicate instruments and two test cycles one nos. each for sterilization and vaccum	Should have minimum five precent sterilization programmes open & closed for metal instruments, liquids & delicate instruments and two test cycles one nos. each for sterilization and vaccum
12) b	Auto shut in case of no water, overheating and earth quake	Auto shut in case of no water, overheating and earth-quake	Auto shut in case of no water, overheating and earth-quake
	New point	20) Inbuilt printer (thermal) with integrated USB.	20) Inbuilt printer (thermal) with integrated USB.
	New point	21) Pre-on mode activation must be available.	21) Pre-on mode activation must be available.
	General Requirements:		

c)	<p>The quoted equipment shall be CE certified (under MDD) along with declaration of conformity or US FDA documentary evidence to that effect shall be uploaded.</p> <p>Bidders are requested to go through the CE / US FDA policy document for submission of CE and US FDA approvals as mentioned elsewhere in tender document.</p>	<p>The quoted equipment shall be CE certified (under MDD) along with declaration of conformity or US FDA documentary evidence to that effect shall be uploaded.</p> <p>US FDA or CE for the distiller only and not for equipment.</p> <p>Bidders are requested to go through the CE / US FDA policy document for submission of CE and US FDA approvals as mentioned elsewhere in tender document.</p>	<p>The quoted equipment shall be CE certified (under MDD) along with declaration of conformity or US FDA documentary evidence to that effect shall be uploaded.</p> <p>US FDA or CE for the distiller only and not for equipment.</p> <p>Bidders are requested to go through the CE / US FDA policy document for submission of CE and US FDA approvals as mentioned elsewhere in tender document.</p>
----	---	---	---

Sd/-
Dy. Ch. E.(CPD)

Sd/-
Dean (KEM Hospital)

Sd/-
Dean (LTMG Hospital)

Sd/-
Dean (BYL Nair Hospital)

Sd/-
Director (ME&MH)

Sd/-
DMC(CPD)